



Republic of the Philippines
 Department of Science and Technology
ADVANCED SCIENCE AND TECHNOLOGY INSTITUTE



ASTI - FM 03-17
 REV 0 / 02 April 2018

PURCHASE ORDER

Supplier:	Rohde & Schwarz Asia Pte. Ltd.	PO No.:	18-12-0341
Address:	9 Changi Business Park Vista, 01-01 Singapore 486041	PO Date:	December 05, 2018
TIN:	430-591-418-000	Mode of Procurement:	NP: Two Failed Biddings

Gentleman:

Please furnish this Office the following articles subject to the terms and conditions contained herein:

**Electronics Product Development Center
 Department of Science and Technology
 Bicutan, Taguig City**

Place of Delivery:

Delivery Term:

on or before 28 February 2019

Date of Delivery:

Payment Term:

Government Terms

Warranty Term:

Stock / Property No.	Unit	Description	Quantity	Unit Cost	Amount
1	lot	<p>EMS (Susceptibility/ Immunity) Measurement Test System</p> <p>Electromagnetic Susceptibility (EMS) TURNKEY SYSTEM</p> <p>- This purchase requisition is for a Turnkey EMS Test System inclusive of design, supply, integrate, install and commission with the existing Electro Magnetic Interference (EMI) Test System. The test system shall be for full compliance, fully automated, controlled by a modular based test system software for testing various electronic and electrical equipment according to latest editions of the following standards:</p> <p>IEC 61000 4-2/ Electrostatic Discharge (ESD) Immunity IEC 61000-4-3/ Radio Frequency (RF) Radiated Immunity IEC 61000-4-4/ Electrical Fast Transients (EFT)/Burst Immunity IEC 61000-4-5/ Surge/Combination Wave Immunity IEC 61000-4-6/ RF Conducted Immunity IEC 61000-4-8/ Power Frequency Magnetic Field Immunity IEC 61000-4-9/ Pulse Frequency Magnetic Field Immunity IEC 61000-4-11/ Voltage Dip, Interruption, Variation Immunity IEC61000-6-1/ Residential, Commercial, light Industrial Immunity IEC61000-6-2/ Industry Immunity ISO 11452-2/ Automotive RF Radiated Immunity ISO 11452-4/ Automotive Bulk Current Injection CISPR 14-2 CISPR 35</p> <p>Bidders must be able to provide a turnkey solution for a complete EMS test setup. Bid price should be inclusive of all costs necessary (permits, taxes and duties, etc.) DDP (Delivered Duty Paid). DDP will be based on the rules of Inco terms 2010.</p> <p>Progressive payment terms: 1. Compliance of EMS Test System Factory Acceptance Test (FAT): 10 percent 2. Supply, delivery and installation of EMS Test Systems: 40 percent 3. Commissioning of EMS Test Systems: 30 percent 4. Completion of EMS Test System Training (SAT): 20 percent</p>	1	₱79,501,832.53	₱79,501,832.53

Although the Bid will be for one (1) lot, we will be requiring that the External Provider still have to indicate the prices of each item individually (for future reference).

After Sales Service & Maintenance:

The External Provider (Principal) shall have a local service & calibration center based in the Philippines for diagnosis, maintenance and repair of test and measurement instruments. They must provide a list of their service centers.

Warranty:

During the Warranty period, the External Provider upon proper notification of the Buyer shall:

- (1) Send representatives within two (2) calendar days to verify and troubleshoot the issue upon notification. Experts shall arrive within a seven (7) calendar days at the site where the defect(s) of Supplies is (are) found to repair or replace;
- (2) Bear all expenses arising from this repair/replacement;
- (3) Must immediately notify the Buyer and/or End User in case the Supplies cannot be repaired on site;
- (4) Notify the Buyer estimated time for repair/replacement but the period for repair/replacement shall not be longer than five (5) weeks from the dispatch date of the defective and/or damaged items;
- (5) Bear all costs arising from or in connection with the repair and/or replacement, including but not limited to customs duties, taxes and transportation fees;
- (6) Bear all costs arising from and/or in connection with the delay in case the items dispatched for repair/replacement cannot be returned within the stipulated time;
- (7) Extend the warranty period in days, from the day the issue has been reported to the day the repair/replacement has been successfully made, after the verification of the issue.

Any replacement or repair provided under warranty period shall be warranted by the External Provider for at least another 3 (three) months on top of the remaining warranty period from the date of finishing repair and/or installation of replaced items. Extended warranty shall only cover the repaired part of the equipment.

Warranty: 3 years

Delivery Date: On or before February 28, 2019.

Duties and Responsibilities of the Supplier:

1. Supply complete turnkey system with the specifications equal to or better than as shown below.
2. Conduct comprehensive training for three (3) key ASTI and EPDC personnel responsible for EMC tests during the Factory Acceptance Test (FAT) for at least two (2) standard measurements and a maximum of six (6) participants from ASTI and EPDC that will oversee the Factory Acceptance Test (FAT). Conduct full comprehensive training for at least 5-10 ASTI and EPDC personnel after installation and commissioning of the Turnkey system for the Site Acceptance Test (SAT).
3. Provide at least two (2) copies of the operations and training manual/documentation for the Turnkey system in printed (hard copy) and electronic (soft copy) formats.
4. Provide technical support and warranty for three (3) years. The Supplier shall render after sales services to the EPDC Project.
5. Ensures that the existing EMI Test System will not be affected (with regards to test integrity) due to the installation.

Responsibilities of ASTI and EPDC Project Team:

- Reject any unit or any part thereof that fail to pass any test and/or inspection or do not conform to specifications.
- The Supplier shall either rectify or replace such rejected unit or parts thereof or make corrections necessary to meet the specifications at no cost to ASTI, and shall repeat the test and/or inspection, at no cost to ASTI.

Others:

- Technical support for Three (3) years (includes aftersales services & maintenance)
- Customer concerns shall have response within two (2) days upon receipt of notification. Expert Technical Assistance within one (1) week.

Technical Specifications:

- The EMS test system should be supplied in rack(s) with fans and internal cabling.
- Calibration Certificates should be supplied as well as Conformance Certificates (as necessary).
- It must include the following accessories apart from equipment:
 - Interlocking system to 10m SAC.
 - Power Distribution unit should be included

A. Radiated Immunity Test System

1) 1 set RF Shielded Amplifier Room:

- Dimensions: Approx. 3.1m x 3.1m x 3.0m (LxWxH)
- Shielding Effectiveness (based on EN 50147-1 or MIL-STD-285)
 - * Shielding Effectiveness Value
 - 100kHz: $\geq 90\text{dB}$ (Magnetic Field)
 - 1MHz: $\geq 100\text{dB}$ (Magnetic Field)
 - 100MHz: $\geq 100\text{dB}$ (Electric Wave)
 - 2GHz: $\geq 100\text{dB}$ (Plane Wave)
 - 10GHz: $\geq 100\text{dB}$ (Micro Wave)
- Construction
 - * Self-supporting structure
 - * Outside & inside shielding shall have a smooth surface.
- Flat Floor with anti-static floor cover
 - * Max. load capacity: 5000 N/m²
- Single leaf, manual door
 - * RF shielded. Door fingers shall not be exposed to prevent damages.
 - * Size: clear opening $\geq 0.9\text{m}$ (w) x 2.0m (h)
- Honeycomb air vents with Waveguide ventilation
 - * Quantity: 4 pcs
- Power line filters: 1- Φ (2-lines) 270 VAC/ 30A (for lighting, emergency light, receptacles, EUT, etc.)
- Power line filters: 3- Φ (4-lines) 480 VAC/ 30A (for power amplifier, etc.)
 - * Attenuation (insertion loss) $\geq 100\text{dB}$ ($150\text{kHz} \leq f \leq 18\text{GHz}$)

2) 1 set RF Power Amplifier:

- General requirement:
 - To fulfil 80MHz – 1GHz, 30V/m @ 3meter distance 80%AM for horizontal & vertical (IEC 61000-4-3:2006 Ed. 3.0)
 - Fulfil 80MHz – 200MHz, 100V/m @ 1m distance for vertical for ISO 11452-2 (automotive)
 - Also fulfil 200MHz – 1GHz, 200V/m @ 1m distance for vertical and horizontal for ISO 11452-2 (automotive)
- Frequency range: 80MHz – 1GHz
- Nominal input impedance: 50 Ω
- Power output
 - * min. 1000W (<400MHz) @ P1dB
 - * min. 850W (>400MHz) @ P1dB
- Gain flatness: $\pm 3.0\text{dB}$
- With built-in or external Directional Coupler
- Remote control via LAN or GPIB

3) 1 set RF Power Amplifier:

- General requirement:
 - To fulfil 1GHz – 6GHz, 30V/m @ 1m distance for vertical and horizontal, 80%AM (IEC 61000-4-3:2006 Ed. 3.0)
 - Also fulfil 1GHz-4GHz, 200V/m @ 1m distance for vertical and horizontal, for ISO 11452-2 (automotive)
- Frequency range: 1GHz - 6GHz
- Nominal input impedance: 50 Ω
- Power output:
 - 1GHz – 2.5GHz: min. 300W @ P1dB
 - 2.5GHz – 6GHz: min. 100W @ P1dB
- Gain flatness/Variation: $\pm 3.0\text{dB}$

- With built-in or external Directional Coupler
 - Remote control via LAN or GPIB
- 4) 1 set Power Meter:
- 4 channels
 - Frequency range: sensors dependent
 - Remote control via LAN or GPIB
 - Can be Rack Mounted with adapter
- 5) 2 units Power sensor:
- One for forward power & one for reverse power
 - Frequency range: 9kHz – 6GHz
 - Power measuring range: -55dBm to +10dBm
 - SWR: $\leq 1.15: 1$ (9kHz to 2GHz)
 - SWR: $\leq 1.35: 1$ (2GHz to 6GHz)
 - With Calibration Certificate from OEM
- 6) 1 set RF Switching Unit:
- General requirement:
 - * Designed for EMS Measurement, switching from single signal source to multiple amplifiers, support multiple paths of power reading, support interlock interface with chamber's door
 - Must have automatic switching capability, controllable via system software and manual front panel control
 - Frequency range: DC – 18GHz
 - 2 x RF relay, SP6T & 4x RF Relay SPDT
 - Remote Control via LAN or GPIB
 - Modular in design, usable up to 40GHz, additional 12 path switching for possible future upgrade
 - Can be Rack Mounted with adapter
- 7) 1 set Log-periodic antenna for IEC 61000-4-3 testing:
- Frequency range: 80MHz to 6000MHz
 - Isotropic Gain @ 80MHz: $\geq 6.7\text{dBi}$
 - VSWR: < 2
 - Max input power:
 - 4000W @ 80MHz
 - 1500W @ 400MHz
 - 950W @ 1000MHz
 - 500W @ 2500MHz
 - 300W @ 6000MHz
 - With complete antenna adapter and movable Stand
 - Automatic Linear polarization with pneumatic and electric remote control
- 8) 1 set Log-periodic antenna for ISO 11452-2 testing:
- Frequency range: 80MHz to 1000MHz
 - Isotropic Gain @ 80MHz: $\geq 6.5\text{dBi}$
 - VSWR: < 1.5
 - Max input power: 2000 W
 - With movable Stand
 - Automatic Linear polarization with pneumatic and electric remote control
- 9) 1 set Horn Antenna for ISO 11452-2 testing:
- Frequency range: 1GHz – 4GHz
 - Gain: minimum 11dBi @ 1GHz
 - VSWR: < 1.6 (typ./average)
 - Max input power: 500W @ 4GHz
 - With complete antenna adapter and stand
- 10) 1 set Field Monitoring Equipment:
- Electric field sensor
 - Isotropic detection (X,Y,Z axis reading)
 - Frequency range: 100kHz – 6GHz
 - Dynamic range: 0.5 to 800V/m
 - Laser power supply
 - With Calibration Certificate from OEM
 - Low dielectric probe stand
 - Fiber-optic link control line
- 11) EUT Low Dielectric Table
- Size: 150cm x 100cm x 80cm (L x W x H)
 - Load capacity: 100kg
 - Polystyrene/Styrofoam material
 - Base and Table Dielectric Constant: < 1.03

- Laminated Top Dielectric Constant: < 3.0
- With EUT placement grid on test table top

B. Conducted Immunity Test System

1) 1 set Power Meter

- 4 channels
- Frequency range: sensors dependent
- Remote control via LAN or GPIB
- Can be Rack Mounted with adapter

2) 3 units Power sensor

- One for forward power, one for reverse power, one for calibration path
- Frequency range: 9kHz – 6GHz
- Power measuring range: -55dBm to +10dBm
- SWR: $\leq 1.15: 1$ (9kHz to 2GHz)
- SWR: $\leq 1.35: 1$ (2GHz to 6GHz)
- With Calibration Certificate from OEM

3) 1 set RF Power Amplifier

- General requirement
 - a. To fulfil 150kHz – 80MHz, 10Vrms 80%AM (IEC 61000-4-6:2006 Ed. 2.2)
 - b. Also covers ISO11452-4 (automotive) 1MHz – 400MHz, 200mA substitution test
- Frequency range: 10kHz – 400MHz
- Power output: ≥ 250 W @ P1dB
- Gain Variation / Flatness: ± 3 dB
- With built-in or external directional coupler
- Remote control via LAN or GPIB

4) 1 set Coupling Decoupling Network (CDN)

- Supply Line
 - a. 1 Line, (a pair)
 - b. 2/3 Lines, Switchable, 16A
 - c. USB connector
 - d. T8S
- With calibration sets

5) 1 set EM Clamp

- Compliant to IEC61000-4-6 Ed. 4 requirement
- Coupling aperture: 20mm
- Input power rating:
 - a. 150kHz to 100MHz: 100W CW
 - b. 100MHz to 230MHz: 100W CW
 - c. 230MHz to 1GHz: 50W CW
- Include calibration adapter
- Decoupling Clamp Included, 20mm coupling aperture

6) 1 set Bulk Current Injection (BCI)

- Compliant to IEC61000-4-6 Ed. 4 requirement
- Current Injection Probe
 - a. Frequency range: 10kHz to 400MHz
 - b. Window diameter: 43mm
 - c. Outside diameter: >112mm
 - d. Max. input power: 1000W
 - e. Calibration jig included
- Monitoring Probe
 - a. Frequency range: 10kHz to 400 MHz
 - b. Insertion loss: Typical approx. -22dB
 - c. Transfer impedance: typical approx. 4 Ω (100kHz to 230MHz)

7) 1 set Oscilloscope

- Input channels: ≥ 2
- Analog bandwidth: ≥ 3 GHz
- Rise time/fall time: ≤ 118 ps (10% to 90% at 50 Ω) or less
- RMS noise floor at 50 Ω (typ.)
 - a. 1 mV/div ≤ 0.16 mV
 - b. 20 mV/div ≤ 0.35 mV
 - c. 50 mV/div ≤ 0.8 mV
 - d. 200 mV/div ≤ 3.40 mV
- Waveform arithmetic/measurement
 - * Peak to peak, minimum, maximum, average, RMS

C. System Accessories

1) 1 set Integrated System Software (Radiated Immunity)

- Functionality – Control the complete EMS test system with drivers for wide selection of measuring instruments and system components such as test receiver, spectrum analyzer, signal generator, antenna mast, turntable and etc.
- Shall be compliance to IEC 61000-4-3 latest standard
- Shall come with latest requirement for amplifier 2dB linearity check function
- Measurement results shall display the results for factor measurement of level monitoring, calibration of injection level, saturation check and immunity measurement
- Shall be able to perform 16-points uniformity test as per IEC 61000-4-3
- Field probe position points shall be able to set from 1 point to 16 points or more during uniformity test
- Software shall support simultaneously 1, 2, 3 and 4 probes for uniformity test
- Uniformity test shall be able to be saved and continued the next day by recalling the same saved file name
- Software shall support existing R&S SMB100A RF Signal Generator
- Software shall support at least 3 different mainstream manufacturer of RF signal generators and power meters for easy replacement during service period
- Shall have function to export uniformity measure result, calibration reference result, linearity check result and immunity test results to excel, words.
- Measurement setup shall include measurement graph, monitoring frequency, SG level, level meter level, field strength level, forward power and reverse power in one window
- Testing measurement condition shall be able to be saved and load from measurement software
- Fail point shall be able to be marked during immunity measurement with remarks and comments
- Amplifier Linearity Check measurement shall be done easily in one window displaying one graph showing frequency, forward power, reverse power, power difference and Total Judgement
- Amplifier Linearity check calibration output difference for upper and lower tolerance shall be able to be easily changed to adapt future changes in standard.
- Software shall have one-point fast verification check function to be used for daily or weekly or monthly check of system stability compare to the reference measurement during uniformity measurement
- Measurement of Immunity frequency shall be able to be set 3 modes :
 - a. step mode (start & stop frequency) with interval in percentage or frequency
 - b. point mode (dedicated frequency point)
 - c. step + point mode for flexible testing
- Modulation of Test level shall be able to be set for AM, FM, Pulse or external signal generator
- Measurement shall be able to be paused during measurement
- Frequency point shall be able to be freely moved during measurement process
- Real time monitoring of forward power and reverse power shall be display on software
- Shall have the ability to set upper & lower limit of forward and reverse power reference to the calibration power to ensure safety of system
- Shall have the ability to set upper limit of signal generator output to protect the system
- Software shall have function to input factor for field probe calibration factor
- Software shall have function to input factor of forward and reverse power of the amplifier
- Operating system compatibility: Latest Windows versions

2) 1 sets Integrated System Software (Conducted Immunity)

- Functionality – Control the complete EMS test system with drivers for wide selection of measuring instruments and system components such as test receiver, spectrum

Analyzer, signal generator, and etc.

- Software shall support at least 3 different mainstream manufacturer of RF signal generators and power meters for easy replacement during service period
- Shall have function of Saturation Check to perform in order to verify that the amplifier is not saturated as per new requirement in IEC 61000-4-6
- Measurement results display to display the results for factor measurement of level monitoring, calibration of injection level, saturation check and immunity measurement
- Shall have function to export factor level monitoring path, calibration result, saturation test result, and immunity test results to excel, words.
- Measurement display shall include measurement graph, monitoring frequency, input level, output level, SG level, level meter level, forward power, reverse power and SWR power in one window
- Testing measurement condition shall be able to be able to be saved and load from measurement software
- Calibration of injection level setting shall be able to factor in level monitor path with additional factor setting
- Saturation Check measurement shall be done easily in one window displaying one graph showing frequency, test level, SG level, monitor level, level meter level, level difference, forward power, reverse power, power difference and Total Judgement
- In saturation check calibration, changes of power difference for upper and lower tolerance shall be able to be easily changed to adapt future changes in standard.
- Judgement of saturation check shall be able to be selected either judge by power difference or judge by monitor level difference
- Measurement of Immunity frequency shall be able to be set 3 modes:
 - a. step mode (start & stop frequency) with interval in percentage or frequency
 - b. point mode (dedicated frequency point)
 - c. step + point mode for flexible testing
- Measurement of Immunity Test level shall be able to be set in Voltage, Voltage (e.m.f.) or current with start and stop test level.
- Modulation of Test level shall be able to be set for AM, FM, Pulse or external signal generator
- Measurement shall be able to be paused during measurement
- Frequency point shall be able to be freely moved during measurement process
- Software shall have real time monitoring of forward power and reverse power display on software
- Software shall have ability to set upper & lower limit of forward and reverse power reference to the calibration power to ensure safety of system
- Software shall have ability to set upper limit of signal generator output to protect the system
- Software shall come with function to compare calibration factors to observe the stability of the measurement of different time
- Shall be working with existing RF Signal Generator (R&S SMB100A)
- Operating system compatibility: Latest Windows version

3) 2 sets Control PC/Laptop

- Intel Core i7 Processor or better
- Memory 8GB (2x4GB) or better
- 1TB 7200RPM S-ATA HDD or better
- Integrated Intel HD Graphics 2000 or better
- Windows 10-64bit or better
- 23" Wide Monitor

4) 1 set Cables

- Optical cabling, RF cabling, GPIB adapter and cable that will be suitable for the application and implementation of the system.

5) 1 set Rack

- 19" Rack and AC power distribution unit suitable to the system

D. Transient Test Requirements

1) 1 set Electrostatic Discharge Tester

- Discharge Voltage:
 - a. 1kV to 30KV (Air)
 - b. 1kV to 30KV (Contact)
- Battery &/or mains powered with display on ESD Gun.
- User replaceable battery feature (Requires AA size rechargeable battery in the ESD pistol and not requiring the disassembly/assembly of the battery pack that will void the warranty of the equipment).
- ESD pulse holding time $\geq 5s$ (IEC 61000-4-2)
- Polarity: Positive/ Negative/ Alternate
- Discharge Network: 150pF/330 Ω (IEC61000-4-2)
- ESD Voltage verification kit with measurement from oscilloscope
- ESD Target with vertical plate as per requirements in IEC 61000-4-2
 - a. Variation of insertion loss: <0.5 dB up to 1GHz
 - b. Frequency range: up to 4GHz
 - c. Input impedance: less than 2.1 ohm
 - d. Input voltage: 30 kV
 - e. Current range: 0 – 120A
 - f. Come with stand supporting ESD Gun during verification

2) 1 set Multi-Transient Generator

- Coupling/Decoupling network at mainframe: AC/DC 300V, 16A
- EUT input with overcurrent protection
- Built-in EUT voltage, current and frequency monitor show on front screen
- Built-in temperature and humidity sensor
- Front panel impulse voltage & current monitor via BNC
- Large touch screen operation
- Front panel input power selection buttons
- IEC 61000-4-4 Electrical Fast Transient(EFT) / Burst
 - a. Burst Voltage: 500V to 4000V (level 1 – 4)
 - b. Burst Frequency: 1kHz to 1000kHz
 - c. Burst time: 0.01ms to 30ms
 - d. Rep. Time: 1ms – 1000ms
- IEC 61000-4-5, Surge (1.2/50us – 8/20us)
 - * General Requirement:
 - a. High energy spikes like a lightning surge can be coupled from the nearby Power line back to the local power grid.

All equipment connected to the power grid will experience the high energy spikes
 - b. Surge Voltage: 500V to 4000V (level 1 – 4)
 - c. Surge Current: 250A to 2000A
 - d. Rep. Time: minimum 1 per 5 seconds @ 4kV
 - e. Phase Syn: 0° to 359°
- IEC 61000-4-11, Dips & Drops
 - a. Voltage Requirement: 0% to 100% Automatic Variable
 - b. Peak In-rush: 500A
 - c. Switching Times: 1 to 5 μs (100 Ω load)
 - d. Event Time: 50 μs to 30s,
 - e. Phase Synchronization: 0° to 359°(1°Step)
 - f. Internal Couplers
 - i) EUT Supply: 16A max, 1- Φ ,
 - ii) EUT Vac: 0 to 250Vrms, 50/60Hz
 - iii) EUT Vdc: 20 to 300Vdc
 - iv) Ethernet control as the system may need to be moveable between SAC and CR.
- High Speed Telecom line Coupler
 - a. Maximum Surge Test Voltage: 4kV (1.2 / 50 μs pulse as per IEC 61000-4-5)
 - b. Common mode coupling to 8 symmetrical lines
 - c. 25 Ω , 40 Ω , 80 Ω , 160 Ω , 320 Ω coupling elements with capacitive coupling elements
 - d. Applications –Ethernet 1Gbps,
 - e. 8 wire unshielded twisted pair (UTP) cable
 - f. EUT Voltage: max 300Vdc
 - g. EUT Current: max 1A
- Capacitive Coupling clamp for IEC61000-4-4 (Dataline and Telecom lines)
 - a. Cable Diameter: Up to 70mm
 - b. Come with verification plate as per IEC 61000-4-4
- IEC 61000-4-8, IEC 61000-4-9, Power and Pulse

Magnetic Field

- a. Power Field Strength: 1 to 1000A/m (level 5)
- b. Pulse Field Strength: up to 1000A/m (level 5)
- c. Minimum coil size: 1m x 1m with stand, 1-Turn coil
- d. Come with Current clamp meter for calibration
- Verification Accessories for EFT & Surge
 - a. Voltage Differential Probe for surge voltage
 - b. Current Probe for surge current
 - c. Calibration adapters 50ohm and 1kohm for EFT
 - d. Calibration adapters for direct pulse output and CDN output for IEC 61000-4-4 & IEC 61000-4-5
- Windows Software for Multi-Generators with reporting capabilities

3) Programmable Power Source

- The AC and DC power sources shall be able to provide three phases, split phase and single phases output modes.
- The output shall be able to provide three phase AC and any single combination phase either AC or DC simultaneously.
- The output voltage shall be 0-310 V ac (Line to Neutral), 0-540 V ac (Line to Line) and 0-425 V dc.
- The AC total rated power output of three phase and single phase shall be ≥ 15 kW
- The DC total rated power out shall be ≥ 15 kW.
- The output voltage waveform shall be able to produce in sine wave, square wave, triangle wave and clipped wave
- The frequency output shall be able to generate at DC and between 15 Hz to 1200 Hz with resolution ≤ 0.01 Hz.
- The AC/DC RMS current in 3 and 2 phase modes shall be ≥ 40 Ampere RMS for AC mode and ≥ 20 Ampere RMS for DC mode.
- The AC/DC RMS current in 1 phase mode shall be up to ≥ 125 Ampere for AC mode and ≥ 62 Ampere for DC mode.
- The current resolution shall be ≤ 0.01 Ampere.
- The current phase angle range shall be at 0 to 359.9 degree.
- The Harmonic Distortion shall be $\leq 0.5\%$ for 50Hz and 60Hz
- The output noise shall be ≤ 150 mV at 50Hz and 60Hz.
- The load regulation at AC mode shall be $\leq 0.05\%$
- The efficiency of the power source shall be $\geq 85\%$.
- The power source shall have measurement function to monitor the voltage, current, power and true power on all phases.
- The weight Programmable power source shall be ≤ 55 kg for portability

E. Transient Testing Room Layout

- External Provider has to provide and do all electrical and grounding installation (ie: outlets, ground plates, etc.) as indicated in the Electrical outlet layout for Transient Testing Room and provide all necessary tables and chairs needed in the Transient Immunity test measurements. This should be supervised by a PRC Licensed Engineer (Civil and Electrical).
- External Provider has to provide a detailed Transient Testing Room Layout that indicates the location of the different test setups in their proposal.
- Provide all necessary peripherals of the test setup in the Transient Testing room for the following tests:
 - a. Harmonics and Flicker setup (Table and chair only)
 - b. Conducted Transient Immunity measurement according to IEC61000-4-4, IEC61000-4-5, IEC61000-4-11
 - c. Magnetic Field Immunity measurement according to IEC61000-4-8, and IEC 61000-4-9
 - d. ESD Simulator Test measurement according to IEC61000-4-2
- Room location: Transient Testing Room
- Note: Please refer to the attached document for the EMC Laboratory Floor Layout and Transient Testing Room Electrical Outlet Layout. (Annex A- EMC Laboratory Floor Layout, 1 page; Annex B- Transient Testing Room Electrical Outlet Layout, 1page).

F. Other Requirements

- 1) EMS Test and Measurement Systems for Factory

Acceptance Test (FAT)

- FAT shall be conducted at an established overseas factory test site. It shall cover the following fields: quantity, marking, origin, technical specifications, operation under applied standards.

- Testing of test instruments would be carried out by External Provider as part of factory acceptance for the following tests:
 - a. Radiated Immunity measurement according to CISPR 14-2, CISPR 35, IEC61000-4-3, IEC61000-6-1, and IEC61000-6-2
 - b. Conducted Immunity measurement according to IEC61000-4-6
 - c. Conducted Transient Immunity measurement according to IEC61000-4-4, IEC61000-4-5, IEC61000-4-11
 - d. Magnetic Field Immunity measurement according to IEC61000-4-8, IEC 61000-4-9
 - e. ESD Simulator Test measurement according to IEC61000-4-2
 - f. Automotive RF Radiated Immunity measurement according to ISO 11452-2
 - g. Automotive Bulk Current Injection measurement according to ISO 11452-4
- A comprehensive training of at least two (2) measurement standards that will be agreed upon by the external provider and the end user. This must be provided by the external provider to at least three (3) representatives of the end user during FAT.
- External Provider shall issue a certificate of completion for the tests and putting into operation for Supplies and System stating that said verification has been performed and that the Supplies' functions have been successfully tested and they can perform in accordance with the Technical specifications as stated in the Technical document.
- Two (2) ASTI Project representatives shall sign on the Certificate.
- If the two (2) ASTI Project representatives, fail to sign the Certificate of Acceptance, the External Provider shall be notified in writing the reasons for the failure.
- A maximum of six (6) participants from ASTI and EPDC, will oversee the Factory Acceptance Test (FAT), pre-delivery inspection and FAT technical training.
- All costs shall be borne by the External Provider.
- Number of Days for FAT: 5 days (Monday-Friday).

2) EMS Test and Measurement Systems for Site Acceptance Test (SAT)

- SAT of measuring instruments shall cover the following fields: quantity, marking, origin, technical specifications, and operation under applied standards.

- Testing of test instruments would be carried out by External Provider as part of SAT for the following tests:
 - a. Radiated Immunity measurement according to CISPR 14-2, CISPR 35, IEC61000-4-3, IEC61000-6-1, and IEC61000-6-2
 - b. Conducted Immunity measurement according to IEC61000-4-6
 - c. Conducted Transient Immunity measurement according to IEC61000-4-4, IEC61000-4-5, IEC61000-4-11
 - d. Magnetic Field immunity measurement according to IEC61000-4-8
 - e. ESD Simulator Test measurement according to IEC61000-4-2
 - f. Automotive RF Radiated Immunity measurement according to ISO 11452-2
 - g. Automotive Bulk Current Injection measurement according to ISO 11452-4
- During SAT, The training must be provided to at least 10 pax.
- External Provider and Representative shall sign on the Certificate.
- If Representative fail to sign the Certificate of Acceptance upon completion of the tests, the representative shall immediately notify the External Provider in writing the reasons for the failure.

- A Full comprehensive training must be provided during SAT.
- The program should be designed to cover the following topics:
 - a. Overview of EMC standards (Generic & Product Standards)
 - b. Overview of Test & Measurement System in EMC Lab.
 - c. CISPR & IEC/EN tests for general electrical/electronic & IT equipment
 - d. EMS Test System components and applications
 - e. Immunity measurements software for automated tests
 - f. Calibration of EMS test system setup
 - g. Demonstration of report generation for calibration of system and test results on CISPR & IEC/EN tests
 - h. Work instruction manuals for testing of the demo EUTs.
 - i. The external provider must provide two (2) demo EUTs per standard measurement for the comprehensive training.
 - j. System maintenance and troubleshooting procedures
- External Provider shall issue a certificate of completion of system training to the representatives at the end of the program.
- Costs for the training (food, venue, training materials, etc.) shall be borne by the External Provider.

- 3) Documentation
- The External Provider is required to provide the following list of documents (in soft and hard copy):
- Technical Literature of instruments listed in offer
 - a. Product Datasheets.
 - b. User Handbook/Operators Manual. (after award of tender)
 - c. Technical Manuals (after award of tender)
 - Plan-view layout of anechoic chamber, shield room, amplifier room, and transient testing room with indication of positioning and dimensions of instruments (racks) and tables, etc.
 - As-built Plans for Amplifier Room and Transient Testing Room Layout
 - List of project references in Asia of similar project scale.
 - Implementation schedule of project, stating clearly milestones and stages. Joint events requiring involvement of the Buyer to be highlighted.
 - Training schedule upon project integration completion (after award of tender).
 - Training should cover the following aspects:
 - a. Overview of EMC standards (Generic & Product Standards)
 - b. Overview of Test & Measurement System in EMC Lab.
 - c. Work instruction manuals for testing of demo EUTs
 - d. System maintenance and troubleshooting procedures

	TOTAL:	₱79,501,832.53
(Total Amount in Words)	Seventy-Nine Million Five Hundred One Thousand Eight Hundred Thirty-Two & 53/100 Pesos	

The contract price is inclusive of taxes and other fees or charges. In case of failure to make the full delivery within the time specified above, a penalty of one-tenth (1/10) of one percent for every day of delay shall be imposed on the undelivered item/s. Once the cumulative amount of liquidated damages reaches ten percent (10%) of the amount of the contract, DOST-ASTI may rescind or terminate the contract, without prejudice to other courses of action and remedies available under the circumstances and in accordance with the provisions of the latest implementing rules and regulations of RA 9184.

Conforme:

Very Truly Yours,




JOEL JOSEPH S. MARCIANO, JR., Ph.D.


JSM

Acting Director

(Signature over Printed Name of Supplier)

(Date)

Fund Cluster:	<u>07</u>	ORS / BURS No.:	<u>07 2086012018-12-000291</u>
Funds Available:	<u>₱79,501,832.53</u>	ORS / BURS Date:	<u>December 5, 2018</u>
		Amount:	<u>₱ 79,501,832.53</u>
 GAY CONCEPCION S. BUGAGAO Accountant III			



05 December 2018

NOTICE TO PROCEED
TWO FAILED BIDDINGS

MR. WILBEN N. CONTI
President & Managing Director
Rohde & Schwarz Asia Pte. Ltd.
9 Changi Business Park Vista,
01-01 Singapore 486041

Dear Mr. Conti,

Notice is hereby given to Rohde & Schwarz Asia Pte. Ltd. that work/delivery may proceed for the following procurement details:

Contract Name	:	<u>Supply and Delivery of EMS (Susceptibility / Immunity) Measurement Test System</u>
Purchase Request No.	:	<u>PDH2-18-08-6327</u>
Purchase / Work Order No.	:	<u>18-12-0341</u>
Total Contract Price	:	<u>₱ 79,501,832.53</u>
(inclusive of taxes, import duties and all other charges or fees)		
Total Contract Price in Words	:	<u>Seventy-Nine Million Five Hundred One Thousand Eight Hundred Thirty-Two & 53/100 Pesos</u>

Upon signing receipt of this Notice, you are responsible for performing the services under the terms and conditions of the Agreement/Purchase Order/Work Order and in accordance with the schedule of requirements/delivery schedule.

You are also hereby required to file a Warranty Security for a minimum period of three (3) months, in the case of Expendable Supplies, or a minimum period of one (1) year, in the case of Non-Expendable Supplies, after acceptance by the Procuring Entity of the delivered supplies. It shall be either retention money in an amount equivalent to at least five percent (5%) of every progress payment, or a special bank guarantee equivalent to at least five percent (5%) of the total contract price. The said amounts shall only be released after the lapse of the warranty period or, in the case of Expendable Supplies, after consumption thereof: Provided, however, That the supplies delivered are free from patent and latent defects and all the conditions imposed under the contract have been fully met.

Note that failure to comply with the above requirements and failure to perform the services under the terms and conditions of the Agreement/Purchase Order/Work Order may constitute grounds for its forfeiture.

Please acknowledge receipt and acceptance of this Notice by signing in the space provided below. There are two (2) copies of this document; you may keep one copy and return the other to the Bids and Awards Committee (BAC) Secretariat of the Advanced Science and Technology Institute. Should you have any questions or clarifications, you may reach us at bac-sec@asti.dost.gov.ph.

Respectfully,


JOEL JOSEPH S. MARCIANO, JR., Ph.D.
Acting Director

DATE OF ISSUANCE:

RECEIVED BY:

WARRANTY SECURITY

(Signature over Printed Name)

Type of Warranty Security: _____

Amount of Warranty Security: ₱ _____

O.R. No.: _____

(Date and Time)